



Simultaneous Measurements of Atmospheric HONO and NO₂ via Absorption Spectroscopy using Tunable Mid-Infrared Continuous-wave Quantum Cascade Lasers

The Harvard community has made this article openly available. [Please share](#) how this access benefits you. Your story matters

Citation	Lee, Ben H., Ezra C. Wood, Mark S. Zahniser, J. Barry McManus, David D. Nelson, Scott C. Herndon, Gregory W. Santoni, Steven C. Wofsy, and J. William Munger. 2011. Simultaneous measurements of atmospheric HONO and NO_2 via absorption spectroscopy using tunable mid-infrared continuous-wave quantum cascade lasers. <i>Applied Physics B: Lasers and Optics</i> 102(2): 471-423.
Published Version	doi:10.1007/s00340-010-4266-5
Citable link	http://nrs.harvard.edu/urn-3:HUL.InstRepos:10085277
Terms of Use	This article was downloaded from Harvard University's DASH repository, and is made available under the terms and conditions applicable to Open Access Policy Articles, as set forth at http://nrs.harvard.edu/urn-3:HUL.InstRepos:dash.current.terms-of-use#OAP

Simultaneous measurements of atmospheric HONO and NO₂ via absorption spectroscopy using tunable mid-infrared continuous-wave quantum cascade lasers

Ben H. Lee¹, Ezra C. Wood², Mark S. Zahniser², J. Barry McManus², David D. Nelson², Scott C. Herndon², Gregory W. Santoni¹, Steven C. Wofsy¹ and J. William Munger¹

¹ Harvard University, School of Engineering and Applied Sciences, Cambridge, Ma, U.S.A.

² Aerodyne Research, Inc. Center for Atmospheric and Environmental Chemistry, Billerica, Ma, U.S.A.

Abstract

Nitrous acid (HONO) is important as a significant source of hydroxyl radical (OH) in the troposphere and as a potent indoor air pollutant. It is thought to be generated in both environments via heterogeneous reactions involving nitrogen dioxide (NO₂). In order to enable fast-response HONO detection suitable for eddy-covariance flux measurements and to provide a direct method that avoids interferences associated with derivatization, we have developed a 2-channel tunable infrared laser differential absorption spectrometer (TILDAS) capable of simultaneous high-frequency measurements of HONO and NO₂. Beams from two mid-infrared continuous-wave mode quantum cascade lasers (cw-QCLs) traverse separate 210 m paths through a multi-pass astigmatic sampling cell at reduced pressure for the direct detection of HONO (1660 cm⁻¹) and NO₂ (1604 cm⁻¹). The resulting one-second detection limits (S/N=3) are 300 and 30 ppt (pmol/mol) for HONO and NO₂, respectively. Our HONO quantification is based on revised line-strengths and peak-positions for *cis*-HONO in the 6-micron spectral region that were derived from laboratory measurements. An essential component of ambient HONO measurements is the inlet system and we demonstrate that heated surfaces and reduced pressure minimize sampling artifacts.

1. Introduction

Atmospheric nitrous acid (HONO) photo-dissociates with a lifetime between 10 and 20 minutes to yield nitric oxide (NO) and hydroxyl radical (OH), the main oxidant in the atmosphere. Nighttime HONO formation and photolysis at sunrise can contribute significantly to early morning photochemistry. Observations of mid-day HONO concentrations above levels expected from photo-stationary balance between HONO, NO and OH suggest a yet unidentified light-dependent production mechanism. These measurements have used various instruments in a wide range of environments [1-3]. Thus, HONO may make a larger contribution to the HO_x (= OH + HO₂) cycle in the lower troposphere than has been accounted for by its nighttime formation. In addition, HONO chemistry may contribute to reactivating deposited nitrogen, which was presumed to be permanently removed from photochemical cycle. HONO is also an indoor air pollutant – both emitted directly from combustion processes and formed on various surfaces – and can react with amines to form carcinogenic compounds [4, 5]. NO₂ is proposed as a precursor to HONO formation via heterogeneous reactions. In order to examine the exchange of HONO between the biosphere and atmosphere and the role of NO₂ in this exchange, we have developed a dual-laser spectrometer to simultaneously measure both gases.

46 Several factors make accurate HONO measurements difficult. It is unstable so certified
47 reference gases for HONO do not exist. Its reactivity and solubility also make it prone to
48 sampling losses and artifacts. Consequently, inter-comparisons between different
49 techniques often exhibit significant discrepancies [2, 6]. Further progress on understanding
50 HONO sources and sinks requires a sensitive and unambiguous measurement method.

51
52 There are many methods to detect HONO, but they do not fully satisfy the need for
53 sensitivity, selectivity and fast time response. Analytical techniques based on derivatization
54 partition gaseous HONO into a liquid and subsequently measure the nitrite ion or its
55 derivative by ion or liquid chromatography, long-path photometry or chemiluminescence [4,
56 7-9]. Although these methods can be very sensitive, the need to scrub HONO into solution
57 may introduce sensitivity to any other gas-phase species that react with the solution [6].
58 Such interferences can be corrected if the chemically active species is quantifiable [10].
59 Furthermore, the need for long extraction integration times (a few minutes) precludes the
60 application of these methods to eddy-covariance flux measurements.


61
62 Absorption spectroscopy directly measures atmospheric trace gases without the need for
63 chemical extraction, with calibrations that are based on constant absorption cross-sections
64 (line-strengths) and specificity that can be confirmed by spectral identification. However,
65 absorption spectroscopic analytical methods tend to be expensive, and for many trace gases
66 the fundamental sensitivity is relatively low, requiring either long absorption paths or
67 increased signal averaging time. Both open-path (differential optical absorption
68 spectroscopy, DOAS) [1, 11, 12] and closed-path (tunable diode laser absorption
69 spectroscopy, TDLAS) [13, 14] systems have been utilized to measure HONO. Based on a
70 prior implementation of nitric acid (HNO_3) and NO_2 TDLAS [15-17], we have developed a
71 dual-channel tunable infrared laser differential absorption spectrometer (TILDAS) using
72 continuous-wave quantum-cascade lasers instead of diode lasers to measure HONO and NO_2 .
73 The advantages of using cw-QC lasers in TILDAS over diode lasers in TDLAS are greater
74 mode stability, higher laser power output and the ability to operate both lasers and detectors
75 near room temperature without the need for cryogenic cooling, which facilitates long-term
76 field measurements. The spectrometer is coupled with sample handling and calibration
77 schemes intended to minimize inlet artifacts and provide quality-assurance that the system
78 is working properly. Section 2 describes the spectrometer design and quantifies
79 performance. Section 3 presents results from the investigation of HONO line-strengths and
80 peak-positions. Section 4 describes the sampling scheme and presents preliminary results
81 from field measurements demonstrating the absence of positive and negative artifacts.

82 83 84 **2. Instrument**

85 The main components of the optical table (Figure 1) include two light sources, a reference
86 cell, multi-pass sampling cell and two detectors. For the light source, the spectrometer uses
87 two thermo-electrically cooled QC lasers (Alpes Lasers) operated in continuous-wave mode
88 that output light in the 6-micron spectral region. The laser light is scanned across a
89 frequency spectrum in time by controlling its temperature, which is coarsely tuned with a
90 Peltier element and finely tuned on a milli-Kelvin scale by providing the lasers with a
91 programmable current ramp using a high compliance current source (ILX Lightwave). The

92 resulting laser frequency scan covers approximately 0.2 cm^{-1} with a resolution of about
93 0.001 cm^{-1} per channel. The instrumental line-widths for both lasers are less than 0.001 cm^{-1}
94 (half-width at half-maximum), which is smaller than Doppler broadened widths. The tuning
95 rates of each laser are determined with a germanium etalon.

96
97




98
99
100

101 **Figure 1.** Optical table of the dual cw-QC laser spectrometer. A = QC laser (1660 cm^{-1} , HONO); B = QC laser
102 (1604 cm^{-1} , NO_2); C = astigmatic multi-pass sampling cell; D = thermo-electrically cooled detector; E = reference
103 cell filled with NO_2 and HONO. The blue and red traces represent paths traveled by the HONO (1660 cm^{-1}) and
104 NO_2 (1604 cm^{-1}) laser light, respectively. The traces for the reference cell and normalization are not shown.
105 The optical table has a footprint of $2 \text{ ft} \times 4 \text{ ft}$ ($0.6 \text{ m} \times 1.2 \text{ m}$).

106
107

108 The two lasers are spatially and temporally multiplexed so that even though light from each
109 laser traverses distinct paths inside and outside the sampling cell, both beams are collected
110 by a single detector at alternate times on the order of 1 ms for each laser. The spectra for NO_2
111 (1604.5 to 1604.7 cm^{-1}) and HONO (1659.5 to 1659.7 cm^{-1}) are repeatedly scanned one
112 after the other at a total rate of about 3 kHz and are subsequently averaged in real-time to
113 improve the signal to noise ratio. Approximately 10% of the duty cycle is dedicated to
114 measuring the detector zero light level when both lasers are off. The spectral fitting software
115 (TDLWintel, also responsible for the laser control) determines the absorbance by performing
116 a non-linear fit according to a set of Voigt line shape functions to the recorded spectra and a
117 low-order polynomial fit to the spectral baseline. Mixing ratios are calculated by accounting
118 for the sample pressure and temperature along with spectral broadening and IR line-
119 strengths archived in the HITRAN database for NO_2 [18] and determined experimentally
120 here for HONO (discussed below). Figure 2 shows transmission spectra of HONO and NO_2
121 observed during calibration gas additions.

122



124

125

126 **Figure 2.** Transmission spectra, averaged over 30-seconds, of (A) HONO and (B) NO₂ at 40 torr. The colored
 127 areas represent the fits to the observed spectra (green dots) according to the known peak-position, line-
 128 strength and recorded pressure and temperature.

129

130

131 Ambient air is sampled through an inlet and transported in tubing (discussed below) to the
 132 sampling cell, which is maintained at a constant reduced pressure to minimize spectral
 133 overlap with other infrared-light absorbing species (especially water), while maintaining
 134 sufficient absorption depths for high sensitivity. Light from each laser enters the multi-pass
 135 cell and reflects between two astigmatic mirrors with multilayer dielectric coatings
 136 (reflectivity ~0.998, LohnStar Optics, Inc.). The mirrors are spaced 0.88 m apart and obtain
 137 238 passes, resulting in a total path-length of 210 m. The light exits the cell through the
 138 entrance coupling-hole and is directed onto a thermo-electrically cooled detector (Vigo). It
 139 should be noted that a liquid-nitrogen cooled HgCdTe detector – which due to its larger active
 140 detector area is less susceptible to aiming changes – could be used to improve overall
 141 stability. However, the costs and measurement interruptions associated with liquid nitrogen
 142 fills for a comparable level of sensitivity with the HgCdTe detectors, make the thermo-
 143 electrically cooled detectors a preferred option.


144

145 The optical table also includes secondary and tertiary light paths, both external to the
 146 sampling cell – derived from the reflections off the front and back surfaces of a transmission
 147 beam splitter. One path is directed through a 7 cm path-length reference cell filled with
 148 gaseous HONO and NO₂. The spectra of this transmitted light is continuously measured by a
 149 second detector and is used to “lock” the lasers to the desired absorption features of HONO
 150 and NO₂, which is necessary when ambient levels of the species of interest are too low to
 151 observe significant absorbance in real-time and also preferred for the routine additions of
 152 zero-air, or ambient air scrubbed of these gases. The third beam and detector could be used
 153 to normalize out power variability associated with the light source, if desired, but has not
 154 been implemented. Details regarding the laser control, optical trace and alignment, data
 155 acquisition system, spectral fitting software, detectors and sampling cell have been
 156 discussed at length previously [17, 19-22].

157

158 Absorbance precisions less than $3 \times 10^{-6} \text{ Hz}^{-1/2}$ ($1.4 \times 10^{-10} \text{ cm}^{-1} \text{ Hz}^{-1/2}$) and $5 \times 10^{-6} \text{ Hz}^{-1/2}$
159 ($1.9 \times 10^{-10} \text{ cm}^{-1} \text{ Hz}^{-1/2}$) are achieved for HONO and NO_2 , respectively. The NO_2 channel is
160 slightly noisier due to the inherent variability of this particular QC device. We achieve the
161 same absorbance precision as that of HONO by utilizing normalization [22]. Normalization
162 with intermittent peak-position locking could be implemented in this dual detector
163 configuration, however, it does not improve HONO sensitivity. Furthermore, at the low
164 HONO mixing ratios expected in rural environments, intermittent line-locking is not
165 adequate to maintain the laser tuning. We sacrifice some NO_2 sensitivity, which is not needed
166 for typical ambient NO_2 levels, to ensure accurate HONO spectroscopy. For measurements at
167 cell pressure of 40 torr, the one-second $1\text{-}\sigma$ precisions for HONO and NO_2 are 100 and 10
168 ppt (6 ppt if absorbance precision observed for HONO is achieved for NO_2), respectively.
169 Long-term stability is limited by optical fringes that change with temperature, causing drifts
170 in the spectral baseline. These effects may be minimized by frequent background spectrum
171 subtractions obtained by flushing the cell with HONO and NO_2 scrubbed air, or zero-air,
172 which is generated by passing ambient air over a heated palladium catalyst. This source of
173 zero-air does not significantly alter the water-vapor mixing ratio, which is critical because
174 there are weakly absorbing H_2O features in both the HONO and NO_2 scans (Figure 2). Time
175 averaging improves the precision of HONO and NO_2 measurements by a factor of 10 or
176 better over an integration time of 30 minutes, which is a typical interval to compute eddy
177 covariance over a forest canopy, with background spectra subtractions conducted once
178 every 5 minutes (Figure 3).

179



180

181

182

183 **Figure 3.** The top two panels show HONO (blue) and NO_2 (red) mixing ratios in ppt (pmol/mol) measured in
184 zero-air at 40 torr. The Allan variance plot on the bottom panel shows the decrease in instrument variance
185 with time averaging for both species. Deviation from pure white or random noise occurs due to slow-moving
186 temperature-driven optical fringes, but is addressed with frequent spectral background subtractions, here
187 conducted every fifth minute for 30 seconds (20 seconds to obtain an average background spectrum and 10
188 seconds of flush time). The y-intercepts on the Allan variance plot representing the 1-second 1σ measurement
189 noise for HONO and NO_2 are 9.4×10^3 and $8.8 \times 10^1 \text{ ppt}^2 \text{ Hz}^{-1}$, or 97 and 9.4 ppt $\text{Hz}^{-1/2}$, respectively.

190
191
192
193
194
195
196
197
198
199
200
201
202
203
204
205
206
207
208
209
210
211
212
213
214
215
216
217
218
219
220
221
222
223
224
225
226
227
228
229
230
231
232
233
234

An optimal spectral region is selected based upon the maximum absorption cross-section and minimum absorbance by other gases that may be present in the sample. We selected the 6-micron region (Figure 2) where lasers, detectors and mirrors were all available allowing simultaneous measurements of HONO and NO₂. There are, however, alternative spectral regions where HONO absorbs mid-infrared light much more strongly, which may improve sensitivity. In particular, HONO absorption lines at 1708.998, 1713.511, 1247.165 and 1273.598 cm⁻¹ may increase precision by a factor of three to five, however, the 8.0 micron region would be costly for NO₂ measurement sensitivity and at this time no appropriate lasers are available in the 5.8 micron region.

3. Line-strengths

The accuracy of the mixing ratios obtained using absorption spectroscopy largely depends on the accuracy of the absorption cross-section or line-strength. For long-lived gases and other well-studied species such as NO₂, the absorption parameters are well characterized and available in publications and in databases such as HITRAN [18]. For HONO, however, there are fewer published studies and greater uncertainty in absolute values, which required us to re-evaluate the HONO spectra.

We determined the amount of infrared (1659.1 to 1660.2 cm⁻¹) light absorbed by the *cis* conformer of HONO while sampling a known amount of total (*cis* + *trans*) HONO at the given constant temperature of 303 K, hence at a constant *cis* to *trans* ratio. This *effective* line-strength of *cis*-HONO was obtained by introducing high levels of gaseous HONO (between 300 and 800 ppb) – generated by passing HCl vapor over powdered NaNO₂ [23] – into the sampling cell at low pressures (< 9 torr) to minimize absorption line overlap while maintaining high signal to noise. In parallel, we quantitatively converted the same HONO source to NO using a heated molybdenum catalyst and quantified NO with a pulsed-mode QC laser (1906.73 cm⁻¹, Hamamatsu Photonics) spectrometer calibrated against a traceable NO standard to determine the absolute HONO concentration. Because NO is relatively inert, it is less susceptible to line-losses and can be accurately calibrated using traceable gas mixtures.

Deviation of the sample from the temperature at which our effective line-strengths were determined will result in a change in the *cis* to *trans* ratio. The effective line-strengths can be corrected knowing the *cis-trans* energy barrier, for which there is a large discrepancy amongst reported values [24-28]. In practice, a constant sample temperature is well maintained by heating the inlet, tubing, the optical table and its protective cover. Furthermore, routine calibration gas additions are conducted in the field to ensure the accuracy of the measurements, as discussed in the next section. Lastly, because the time required for isomerization to occur is much shorter (~10⁻¹² seconds) than that needed for ambient air to travel through the inlet and subsequent tubing to the sampling cell (~1 second), we expect the HONO conformers to be in thermal equilibrium and independent of changing ambient conditions.

Line positions and effective line-strengths used in the spectral fits in Figure 2 are listed in Table 1. Additional details regarding this experiment, along with comparison of the relative absorption strengths between *cis* and *trans* conformers, determination of the pressure-broadening coefficient, comparison to values in the ATMOS database and to a high-resolution FTIR spectrum are part of an ongoing analysis that is not yet complete.

Table 1. Peak-position and effective line-strength of *cis*-HONO between 1659.5 and 1659.7 cm⁻¹ measured at 303 K.

Peak position (cm ⁻¹)	Effective line-strength (cm ² molecule ⁻¹ cm ⁻¹) × 10 ⁻²¹
1659.5031	9.571
1659.5099	8.810
1659.5887	10.15
1659.5968	3.283
1659.6238	1.863
1659.6770	12.92
1659.6886	7.419

4. Sampling technique


Figure 4 shows a schematic of the inlet manifold we use for minimizing contact between sampled ambient air and moist surfaces, for removing coarse particles by inertial separation and for routine additions of reference gases and zero-air. HONO (Henry's Law constant, $K_H = 50 \text{ M atm}^{-1}$) – though not nearly as soluble as HNO_3 ($K_H = 2.1 \times 10^5 \text{ M atm}^{-1}$) – is still photochemically active, relatively soluble compared to NO ($1.4 \times 10^{-3} \text{ M atm}^{-1}$) and NO_2 ($1.2 \times 10^{-2} \text{ M atm}^{-1}$), and mostly dissociated above pH 3.3 [29-32]. In addition to the high probability of HONO equilibrating with accumulated particles and moist surfaces, there is strong evidence for heterogeneous reactions converting precursors to HONO [33]. Consequently, both positive and negative HONO artifacts are a concern, leading us to minimize opportunities for the sample to contact humid surfaces and aerosol that would accumulate on a particle filter.

Surface-adsorbed water is minimized by 1) using a siloxyl-coated quartz inlet to make its surface hydrophobic, 2) heating the inlet, the downstream tubing and sampling cell and 3) reducing the pressure by drawing the sample through a critical orifice built into the inlet (Figure 4). The divergent flows after the orifice – with exhaust vented in line with the incoming flow and the sample air forced to make a 180° turn – cause particles with diameters 4 μm or larger to be separated out of the sample flow by inertia [34, 35]. The inlet and subsequent tubing are shielded from light to avoid photolytic losses and photo-enhanced reactions involving surface adsorbed nitrate ions leading to HONO production [33]. A similarly designed inlet was used successfully to measure formaldehyde, formic acid and ammonia, which are also highly surface active [36, 37].

Frequent in-field artifact testing is an integral part of the instrument system. The manifold is equipped with ports at the entrance of the inlet to allow for routine additions of standards

273 and zero-air (Figure 4). An outlet adjacent to the flow-restricting orifice is used to draw a
 274 subsample of the ambient matrix to a heated Mo catalyst to convert all reducible nitrogen
 275 oxides to NO, followed by quantification of the NO by O₃-chemilumiscence. HONO generated
 276 from the HCl + NaNO₂ source, which typically is > 97% pure, is dynamically mixed with
 277 zero-air, added at the inlet entrance in excess of the total sample flow-rate and sampled by
 278 both instruments, thereby providing an independent check on the measurements by tying
 279 the observations to traceable NO standards. A similar test is applied for NO₂, which is
 280 supplied from either compressed gas standards or a permeation tube.

281
282




283
284
285
286
287
288
289

Figure 4. Schematic of the quartz inlet manifold. The inlet and subsequent tubing are shielded from light to prevent photolysis and photo-induced surface reactions.


290 Standard gas additions can also test for temporary surface adsorption that attenuates
 291 atmospheric variations. The rise and fall in absorbance when a standard is switched on or
 292 off should be nearly instantaneous, with only some delay associated with the flushing time
 293 defined by the pumping speed through the inlet, tubing and sampling cell as well as
 294 smearing due to mixing and diffusion. These response times can be determined by fitting
 295 the mixing ratio time series during standard gas additions to a single exponential curve. A
 296 slower response of HONO compared to that of an inert gas would suggest some uptake of
 297 HONO on the surfaces. For the current instrumental configuration, HONO response is
 298 compared to that of NO₂, which also is generally not attenuated by wall interactions [15-17].
 299 A prototype of this inlet system was deployed at the Study of Houston Atmospheric Radical
 300 Precursor (SHARP) campaign in April and May 2009, during which polluted, humid air was
 301 sampled through the inlet and 40 ft (12.2 m) of 3/8" O.D. (9.5 mm) PFA Teflon tubing. Figure
 302 5 shows indistinguishable response times between HONO and NO₂ during additions of both
 303 gases at SHARP.

304

305 The possibility for positive artifacts from the inlet, tubing or cell surface reactions forming
306 HONO is checked by introducing NO₂ into the inlet in addition to the ambient air matrix. An
307 increase in HONO mixing ratios during these additions would indicate reactions involving
308 NO₂ yielding HONO. During the SHARP campaign, we did not observe any artifact HONO
309 when NO₂ was added to ambient air (Figure 6). Even at NO₂ levels exceeding 100 ppb – well
310 above the range of observed values – there was no change in the HONO mixing ratio,
311 demonstrating freedom from positive artifacts based on NO₂.
312
313



317 **Figure 5.** One-second mixing ratios of HONO and NO₂ during calibration gas additions, observed during the
318 Study of Houston Atmospheric Radical Precursor (SHARP) campaign in April and May of 2009. The response
319 times (τ) – determined by the pumping speed and cell volume (5 L) – of HONO and NO₂ are
320 indistinguishable, indicating no preferential loss of HONO through the inlet, tubing and sampling cell.
321
322
323




328 **Figure 6.** One-second mixing ratios of NO₂ (top) and ten-second average mixing ratios of HONO (bottom)
329 during a high-concentration NO₂ addition through 40-feet (12.2 m) of unheated tubing at ambient pressure,
330 conditions under which surface reactions are favored. Note that there is no HONO formed during or
331 immediately following high levels of NO₂ under humid conditions (no positive artifact). The error bars
332 represent the standard deviation of 10-second averaged data.

333
334

335 Figure 7 shows the mixing ratios of HONO and NO₂ measured during the SHARP campaign.
336 The levels of the two species typically co-vary in time as a result of common or co-located
337 emission sources and mixing in the atmosphere. Higher levels are usually observed in the
338 early morning and in the evening when weaker vertical mixing and shallow boundary layer
339 heights promote accumulation of pollutants and photo-chemical losses are reduced
340 compared to mid-day conditions.

341
342



343
344

345 **Figure 7.** Observed 30-minute averaged mixing ratios in ppb (nmol/mol) of HONO (A) and NO₂ (B) during the
346 entire SHARP campaign. Gaps in the data are due to maintenance operations, inlet attenuation/artifact tests
347 and instrument shutdown from power failures.

348
349
350

351 5. Conclusions

352 This dual continuous-wave mode quantum cascade laser spectrometer has achieved one-
353 second detection limits (S/N = 3) for HONO and NO₂ of 300 and 30 ppt, respectively. Spectral
354 averaging with frequent background subtractions allows further reduction in signal noise,
355 improving the HONO and NO₂ detection limits (S/N = 3) to 35 and 3 ppt, respectively, over a
356 30-minute integration period. The detection limit for HONO is higher than that for NO₂ –
357 despite better absorbance precision on this channel – due to much weaker line-strengths for
358 *cis*-HONO between 1659 and 1660 cm⁻¹ compared to NO₂ at 1604 cm⁻¹. A sample-handling
359 scheme that minimizes adsorbed water on the inlet and subsequent tubing that bring
360 sample into the optical cell is effective at preventing both positive and negative HONO
361 artifacts. The system is designed to allow inlet checks to be a part of routine field operation
362 and provide frequent quality checks on the measurement.

363

364 The instrument described in this study was adapted from a previously deployed instrument
365 with a large optical table and near 1 m base-length multi-pass cell. With insulation and a
366 protective cover, the instrument is 3 ft × 6 ft (0.9 m × 1.8 m). The performance demonstrated
367 by this implementation points to the possibility of using a re-designed astigmatic multi-pass
368 cell with a base-length of 47.5 cm and 200 m of absorption path-length. This cell fits on an
369 optical table measuring 43 × 65 cm. Its smaller volume of 1 L allows faster response times
370 and will be more easily portable for field use without any sacrifice in detection limit [38].
371

372 The chief advantages of this dual cw-QC TILDAS spectrometer over previous diode laser and
373 pulsed-mode QC laser systems are higher power output and narrower laser line-widths,
374 which provide improved precision and enable thermo-electrically cooled detectors to be
375 used instead of cryogenically cooled detectors that present logistical challenges for long-
376 term, remote operation in the field. This spectrometer has been deployed at a rural forest to
377 measure the diurnal and seasonal trends in the exchange of HONO and NO₂ between the
378 biosphere and atmosphere.
379

380 381 **Acknowledgments**

382 The authors gratefully acknowledge the assistance of Ryan McGovern, Stanley Huang and
383 Daniel Glen of Aerodyne Research, Inc., Josh McLaren and Bruce Daube of Harvard
384 University, and the entire SHARP research team. This work was supported by the National
385 Science Foundation Awards No. AGS – 0813617 and 0814202. Additional funding for the
386 SHARP campaign was provided by the Houston Advanced Research Center Grant No. H113.
387

388 389 **References**

- 391 1. Alicke, B., et al, *OH formation by HONO photolysis during the BERLIOZ experiment*. Journal of
392 Geophysical Research-Atmospheres, 2003. **108**(D4): p. -.
- 393 2. Liao, W., et al, *Observations of HONO by laser-induced fluorescence at the South Pole during ANTCTI 2003*.
394 Geophysical Research Letters, 2006. **33**(9): p. -.
- 395 3. Zhou, X.L., et al, *Summertime observations of HONO, HCHO, and O-3 at the summit of Whiteface*
396 *Mountain, New York*. Journal of Geophysical Research-Atmospheres, 2007. **112**(D8): p. -.
- 397 4. Park, S.S., et al, *Investigation of nitrous acid concentration in an indoor environment using an in-situ*
398 *monitoring system*. Atmospheric Environment, 2008. **42**(27): p. 6586-6596.
- 399 5. Sleiman, M., et al, *Formation of carcinogens indoors by surface-mediated reactions of nicotine with*
400 *nitrous acid, leading to potential thirdhand smoke hazards*. Proceedings of the National Academy of
401 Sciences of the United States of America, 2010. **107**(15): p. 6576-6581.
- 402 6. Kleffmann, J., et al, *Intercomparison of the DOAS and LOPAP techniques for the detection of nitrous acid*
403 *(HONO)*. Atmospheric Environment, 2006. **40**(20): p. 3640-3652.
- 404 7. Dibb, J.E., et al, *Soluble reactive nitrogen oxides at South Pole during ISCAT 2000*. Atmospheric
405 Environment, 2004. **38**(32): p. 5399-5409.
- 406 8. Heland, J., et al, *A new instrument to measure gaseous nitrous acid (HONO) in the atmosphere*.
407 Environmental Science & Technology, 2001. **35**(15): p. 3207-3212.
- 408 9. Zhou, X.L., et al, *A method for the measurement of atmospheric HONO based on DNPH derivatization*
409 *and HPLC analysis*. Environmental Science & Technology, 1999. **33**(20): p. 3672-3679.
- 410 10. Kleffmann, J. and P. Wiesen, *Technical Note: Quantification of interferences of wet chemical HONO*
411 *LOPAP measurements under simulated polar conditions*. Atmospheric Chemistry and Physics, 2008. **8**: p.
412 6813-6822.

- 413 11. Kurtenbach, R., et al, *Investigations of emissions and heterogeneous formation of HONO in a road traffic*
414 *tunnel*. Atmospheric Environment, 2001. **35**(20): p. 3385-3394.
- 415 12. Platt, U., et al, *Observations of Nitrous-Acid in an Urban Atmosphere by Differential Optical-Absorption*.
416 Nature, 1980. **285**(5763): p. 312-314.
- 417 13. Li, Y.Q., J.J. Schwab, and K.L. Demerjian, *Fast time response measurements of gaseous nitrous acid using*
418 *a tunable diode laser absorption spectrometer: HONO emission source from vehicle exhausts*. Geophysical
419 Research Letters, 2008. **35**(4): p. -.
- 420 14. Schiller, C.L., et al, *Atmospheric measurements of HONO by tunable diode laser absorption spectroscopy*.
421 Journal of Atmospheric Chemistry, 2001. **40**(3): p. 275-293.
- 422 15. Horii, C.V., et al, *Fluxes of nitrogen oxides over a temperate deciduous forest*. Journal of Geophysical
423 Research-Atmospheres, 2004. **109**(D8): p. -.
- 424 16. Horii, C.V., et al, *Atmospheric reactive nitrogen concentration and flux budgets at a Northeastern US*
425 *forest site*. Agricultural and Forest Meteorology, 2005. **133**(1-4): p. 210-225.
- 426 17. Horii, C.V., et al, *Nitric Acid and Nitrogen Dioxide Flux Measurements: a new Application of Tunable*
427 *Diode Laser Absorption Spectroscopy*. Proceedings of SPIE, 1999. **3758**: p. 152-161.
- 428 18. Rothman, L.S., et al, *The HITRAN 2004 molecular spectroscopic database*. Journal of Quantitative
429 Spectroscopy & Radiative Transfer, 2005. **96**(2): p. 139-204.
- 430 19. McManus, J.B., *Paraxial matrix description of astigmatic and cylindrical mirror resonators with twisted*
431 *axes for laser spectroscopy*. Applied Optics, 2007. **46**(4): p. 472-482.
- 432 20. McManus, J.B., et al, *Comparison of cw and pulsed operation with a TE-cooled quantum cascade infrared*
433 *laser for detection of nitric oxide at 1900 cm(-1)*. Applied Physics B-Lasers and Optics, 2006. **85**(2-3): p.
434 235-241.
- 435 21. Nelson, D.D., et al, *Sub-part-per-billion detection of nitric oxide in air using a thermoelectrically cooled*
436 *mid-infrared quantum cascade laser spectrometer*. Applied Physics B-Lasers and Optics, 2002. **75**(2-3):
437 p. 343-350.
- 438 22. Zahniser, M.S., et al, *Infrared QC laser applications to field measurements of atmospheric trace gas*
439 *sources and sinks in environmental research: enhanced capabilities using continuous wave QCLs*.
440 Proceedings of SPIE, 2009. **7222, 72220H**.
- 441 23. Febo, A., et al, *Evaluation of a High-Purity and High-Stability Continuous Generation System for Nitrous-*
442 *Acid*. Environmental Science & Technology, 1995. **29**(9): p. 2390-2395.
- 443 24. Bongartz, A., et al, *Near-Uv Absorption Cross-Sections and Trans Cis Equilibrium of Nitrous-Acid*. Journal
444 of Physical Chemistry, 1991. **95**(3): p. 1076-1082.
- 445 25. Jones, L.H., R.M. Badger, and G.E. Moore, *The Infrared Spectrum and the Structure of Gaseous Nitrous*
446 *Acid*. Journal of Chemical Physics, 1951. **19**(12): p. 1599-1604.
- 447 26. Mcgraw, G.E., D.L. Bernitt, and Hisatsun.Ic, *Infrared Spectra of Isotopic Nitrous Acids*. Journal of
448 Chemical Physics, 1966. **45**(5): p. 1392-&.
- 449 27. Sironneau, V., et al, *Absolute line intensities of HONO and DONO in the far-infrared and re-*
450 *determination of the energy difference between the trans- and cis-species of nitrous acid*. Journal of
451 Molecular Spectroscopy, 2010. **259**(2): p. 100-104.
- 452 28. Varma, R. and R.F. Curl, *Study of N2o3-H2o-Hno2 Equilibrium by Intensity Measurements in Microwave*
453 *Spectroscopy*. Journal of Physical Chemistry, 1976. **80**(4): p. 402-409.
- 454 29. Becker, K.H., et al, *Solubility of nitrous acid (HONO) in sulfuric acid solutions*. Journal of Physical
455 Chemistry, 1996. **100**(36): p. 14984-14990.
- 456 30. Chameides, W.L., *The Photochemistry of a Remote Marine Stratiform Cloud*. Journal of Geophysical
457 Research-Atmospheres, 1984. **89**(Nd3): p. 4739-4755.
- 458 31. Goretski, J., O.C. Zafiriou, and T.C. Hollocher, *Steady-State Nitric-Oxide Concentrations during*
459 *Denitrification*. Journal of Biological Chemistry, 1990. **265**(20): p. 11535-11538.
- 460 32. Lelieveld, J. and P.J. Crutzen, *The Role of Clouds in Tropospheric Photochemistry*. Journal of Atmospheric
461 Chemistry, 1991. **12**(3): p. 229-267.
- 462 33. Zhou, X.L., et al, *Photochemical production of nitrous acid on glass sample manifold surface*. Geophysical
463 Research Letters, 2002. **29**(14): p. -.
- 464 34. Loo, B.W. and C.P. Cork, *Development of High-Efficiency Virtual Impactors*. Aerosol Science and
465 Technology, 1988. **9**(3): p. 167-176.
- 466 35. Marple, V.A. and C.M. Chien, *Virtual Impactors - a Theoretical-Study*. Environmental Science &
467 Technology, 1980. **14**(8): p. 976-985.

- 468 36. Ellis, R.A., et al, *Characterizing a Quantum Cascade Tunable Infrared Laser Differential Absorption*
469 *Spectrometer (QC-TILDAS) for measurements of atmospheric ammonia*. Atmospheric Measurement
470 Techniques, 2010. **3**: p. 397-406.
- 471 37. Herndon, S.C., et al, *Airborne measurements of HCHO and HCOOH during the New England Air Quality*
472 *Study 2004 using a pulsed quantum cascade laser spectrometer*. Journal of Geophysical Research-
473 Atmospheres, 2007. **112**(D10): p. -.
- 474 38. McManus, J.B., et al, *Application of quantum cascade lasers to high precision atmospheric trace gas*
475 *measurements*. Optical Engineering, 2010. **49**(11).
- 476
- 477
- 478